

## CALIBRATION REPORT: ORDER NO.

JUNE 12, 2020

PAGE 1 OF 1

MANUFACTURER:

OHM-LABS

CS-100

DESCRIPTION:

**CURRENT SHUNT** 

MODEL:

SERIAL:

PROCEDURE:

PROCEDURE:

CALIBRATION DATE:

CS CAL

LAB ENVIRONMENT:

23.4 °C / 44 %RH

12/JUN/2020

MEASUREMENT DATA - AS FOUND & AS LEFT				
APPLIED CURRENT	MEASURED VALUE	UNCERTAINTY		
20 A	1.000 020 3 mΩ	3.0 μΩ/Ω		
40	1.000 020 3	3.0		
60	1.000 020 3	3.0		
80	1.000 017 3	3.0		
100	1.000 010 7	3.1		

## NOTES:

SHUNT WAS ALLOWED TO FULLY STABILIZE AT EACH APPLIED CURRENT.

## STANDARDS USED

ID	DESCRIPTION	MAKE & MODEL	CAL DUE
AS3001	RESISTANCE STANDARD	OHM-LABS 200	31/MAR/2021
AS3401	RESISTANCE BRIDGE	GUILDLINE 9920	28/FEB/2021

## COMMENTS:

OHM-LABS, INC. CERTIFIES THAT THIS CALIBRATION IS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST), OR ANOTHER RECOGNIZED NATIONAL MEASUREMENT INSTITUTE, OR DERIVED BY A RATIO TYPE SELF-CALIBRATION TECHNIQUE, AND IS ACCREDITED TO ISO/IEC 17025:2005. OHM-LABS' QUALITY CONTROL SYSTEM MEETS THE REQUIREMENTS OF ANSI/NCSL Z540-1-1994. THE REPORTED UNCERTAINTIES REPRESENT EXPANDED UNCERTAINTIES EXPRESSED AT A CONFIDENCE LEVEL OF APPROXIMATELY 95%, USING A COVERAGE FACTOR OF K=2. THIS UNCERTAINTY IS AT THE TIME OF TEST ONLY AND DOES NOT TAKE INTO ACCOUNT TRANSIT, USAGE, DRIFT OVER TIME, OR OTHER FACTORS AFFECTING STABILITY. THIS DOCUMENT RELATES ONLY TO THE ITEMS IDENTIFIED HEREIN, AND IS IN COMPLIANCE WITH ALL REQUIREMENTS OF THE ABOVE REFERENCED PURCHASE ORDER. THE CALIBRATION PERFORMED WAS IN ACCORDANCE WITH THE CURRENT REVISION LEVEL OF OHM-LABS' QUALITY CONTROL SYSTEM. TRAINED AND QUALIFIED PERSONNEL PERFORMED THE CALIBRATIONS IN ACCORDANCE WITH THE REQUIREMENTS OF ISO/IEC 17025:2005. THIS CERTIFICATE SHALL NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT WRITTEN PERMISSION OF OHM-LABS, INC.

PERFORMED BY

REVIEWED BY:

